DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

December 5, 2014

STRECK, INC.
C/O MS DEBORAH KIPP
REGULATORY AFFAIRS MANAGER
7002 SOUTH 109TH STREET
LA VISTA, NE 68128

Re: k141957
Trade/Device Name: XN CHECK™ BF
Regulation Number: 21 CFR 864.8625
Regulation Name: Hematology quality control mixture
Regulatory Class: II
Product Code: JPK
Dated: October 24, 2014
Received: October 27, 2014

Dear Ms. Kipp:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulations (21 CFR Parts 801 and 809), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Maria M. Chan -S

Maria M. Chan, Ph.D.
Director
Division of Immunology and Hematology Devices
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

510(k) Number *(if known)*

k141957

Device Name
XN CHECK BF

Indications for Use *(Describe)*

XN CHECK BF is used for control and calibration verification of Sysmex XN series (XN-10, XN-11, XN-20, XN-21) analyzers. It is not, however, intended for actual calibration of these analyzers. Assayed parameters include:

WBC-BF (10^3/µl), RBC-BF (10^6/ µl), MN# (10^3/µl), PMN# (10^3/µl), MN% (%), PMN% (%), TC-BF# (10^3/µl)

Type of Use *(Select one or both, as applicable)*

☒ Prescription Use (Part 21 CFR 801 Subpart D) ☐ Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) *(Signature)*

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

510(k) Submitter: Streck
7002 South 109th Street
La Vista, NE 68128

Official Correspondent: Deborah Kipp, Regulatory Affairs Manager
Address: 7002 South 109th Street; La Vista, NE 68128
Phone: 402-537-5215
Fax: 402-537-5317

Date Prepared: April 18, 2014

Names

Trade Name: XN CHECK™ BF
Common Name: Assayed Hematology Control
Classification Name: Hematology quality control mixture (864.8625)
Product Code: JPK
Panel: Hematology

Predicate Device:

XN CHECK™ BF (K120744)

Intended Use:

XN CHECK BF is used for control and calibration verification of Sysmex XN series (XN-10, XN-11, XN-20 XN-21) analyzers. It is not, however, intended for actual calibration of these analyzers. Assayed parameters include:

WBC-BF (10^3/µl), RBC-BF (10^6/ µl), MN# (10^3/µl), PMN# (10^3/µl), MN% (%), PMN% (%), TC-BF# ($10^{3/}$µl)

Description:

XN CHECK™ BF is an in-vitro diagnostic product that contains the following: stabilized red blood cell component(s) and stabilized white blood cell component(s) in a preservative medium. The product is packaged in polypropylene plastic vials with screw caps with a 3 ml fill. The vials will be packaged in (4) welled vacuum formed clamshell container with the Instructions for Use / assay sheet. The product storage conditions are at 2 - 8° C.
Comparison to Predicate Device:

	XN-CHECK BF (K120744)-Predicate Device	**XN CHECK™ BF -Candidate Device**	**Same or Differences**
Intended Use Statement	XN CHECK BF is used for control and calibration verification of Sysmex XN series (XN-10, XN-20) analyzers. It is not, however, intended for actual calibration of these analyzers. Assayed parameters include: WBC-BF (10^3/µl), RBC-BF (10^6/µl), MN# (10^3/µl), PMN# (10^3/µl), MN% (%), PMN% (%), TC-BF# (10^3/µl)	XN CHECK BF is used for control and calibration verification of Sysmex XN series (XN-10, XN-11, XN-20, XN-21) analyzers. It is not, however, intended for actual calibration of these analyzers. Assayed parameters include: WBC-BF (10^3/µl), RBC-BF (10^6/ µl), MN# (10^3/µl), PMN# (10^3/µl), MN% (%), PMN% (%), TC-BF# (10^3/µl)	Addition of the XN-11 and XN-21 analyzers.
Open Vial Stability	30 days	30 days	Same
Closed Vial Stability	84 days	84 days	Same
Reagents	XN CHECK BF contains the following: stabilized red blood cell component(s) and stabilized white blood cell component(s) in a preservative medium.	XN CHECK BF contains the following: stabilized red blood cell component(s) and stabilized white blood cell component(s) in a preservative medium.	Same
Storage Conditions	2 - 8˚C	2 - 8˚C	Same

Discussion of Tests and Test Results:

To substantiate the product performance claims for XN CHECK BF, Streck collected product performance data for the following studies: Open-Vial Stability, Closed-Vial Stability, and Precision Performance. The resultant data set established that XN CHECK BF is safe and effective for its intended use and that the product is stable for the entire product dating. The product fulfills its intended use as instructed in the Instructions for Use.

Conclusions Drawn From Tests:

Study results show XN CHECK BF to be consistently reproducible, substantially equivalent to the predicate products, and stable for the entire product dating. XN CHECK BF is a safe and effective product, which fulfills its intended use when used as instructed in the Instructions for Use.